

Accessories

- **Instrument utility cart** functionally designed general purpose laboratory cart. Supports any PerkinElmer benchtop system.
- **Cassette carrying basket** to accept ten large vial or 12 small vial cassettes. Designed for easy stacking.
- **See the Equipment, Chemicals & Supplies section** in the PerkinElmer Life Sciences catalog.

Physical Data

Dimensions:

Height:	18.5 in. (47 cm)
Width:	40.5 in. (103 cm)
Depth:	32 in. (81 cm)

Weight (Instrument Only):

523 lb (238 kg)
Shipping weight: approximately 700 lb (318 kg)

Electrical Requirements:

117 Vac + 10%, 50/60 Hz, 20 amp protection
 220 Vac + 10%, 50/60 Hz, 10 amp protection
 3-prong grounded plug, 500 watts
 Power consumption: 1150 VA

Environmental:

Operating ambient temperature 15 to 35°C (59 to 90°F)
 Operating relative humidity 30% to 85%

Typical Performance Data

(As measured in factory at Downers Grove, Illinois)

Energy Range: 0-2,000 keV

Efficiency, Normal Count Mode:

		Minimum Acceptable
³ H	0-18.6 keV	60%
¹⁴ C	0-156 keV	95%

Figure of Merit (E²/B), Normal Count Mode:

³ H	1-18.6 keV	400
¹⁴ C	4-156 keV	1,000

Figure of Merit (E²/B), Super Low Level Count Mode:

³ H	1-12.5 keV	880
¹⁴ C	14.5-97.5 keV	6,000

Observed ³H Performance with 11 mL H₂O in 9 mL Ultima Gold LLT and Plastic Vials:

$$E^2V^2/B > 41,000$$

Observed ¹⁴C Performance for Benzene Samples in Glass Vials with Pico-SL Vial Holders:

$$E^2V^2/B > 325,000$$

Note: The efficiencies, backgrounds, and E²/B values for the Normal Count Mode are determined using NIST traceable PerkinElmer sealed large glass vial standards set P.N. 6008500. Super Low Level Count Mode values are determined using NIST traceable PerkinElmer low level sealed large glass vial standards set P.N. 6018914. No maximum is specified for background.



Worldwide Headquarters: PerkinElmer Life Sciences, Inc., 549 Albany Street, Boston, MA 02118-2512 USA (800) 551-2121

European Headquarters: PerkinElmer Life Sciences, Imperiastraat 8, B-1930 Zaventem Belgium

Technical Support: in Europe: techsupport.europe@perkinelmer.com in US and Rest of World: techsupport@perkinelmer.com

Belgium: Tel: 0800 94 540 • **France:** Tel: 0800 90 77 62 • **Netherlands:** Tel: 0800 02 23 042 • **Germany:** Tel: 0800 1 81 00 32 • **United Kingdom:** Tel: 0800 89 60 46

Switzerland: Tel: 0800 55 50 27 • **Italy:** Tel: 800 79 03 10 • **Sweden:** Tel: 020 79 07 35 • **Norway:** Tel: 800 11 947 • **Denmark:** Tel: 80 88 3477 • **Spain:** Tel: 900 973 255

Windows and Windows NT are registered trademarks of Microsoft Corporation in the United States and/or other countries. SPA: Scintillation Proximity Assay Technology is covered by US Patent number 4568649, European patent number 0154734, and Japanese patent application number 84/52452, Nycomed Amersham plc.

All other trademarks or registered trademarks are the property of PerkinElmer Life Sciences, Inc.

© 2002 PerkinElmer Life Sciences, Inc.

P10980

www.perkinelmer.com/lifesciences

Drug Discovery

Research

Clinical Screening

Liquid Scintillation

Tri-Carb® 3170TR/SL

Super Low Level Liquid Scintillation Counter

Description

The Tri-Carb® 3170TR/SL liquid scintillation counter is a computer-controlled benchtop liquid scintillation analyzer, specially configured for detection of extremely low level alpha and beta radioactivity.

Standard Instrument Features

- **QuantaSmart™ software for the Windows NT® operating system** is a 32 bit operating system software that provides a robust multitasking and easy networking environment and unlimited assays in a secure multiuser environment.
- **Hypertext on-line comprehensive HELP** provides context specific help for every screen and data entry field environment.
- **Powerful integrated computer system control** with 128 MB RAM and 20 GB hard disk (minimum) with a 17 inch SVGA monitor, built-in CD ROM and 3.5 inch floppy disk drive.
- **Robust downloading sample changer mechanism** with a double light sealing shutter, allows the photomultiplier tube detectors to remain on for maximum stability even during sample changing.
- **Patented TR-LSC® (Time-Resolved Liquid Scintillation Counting)** provides reduced background and increased sensitivity using unique afterpulse rejection technology.
- **Special, proprietary "Surround TR-LSC" BGO (Bismuth Germanium Oxide) detector guard** surrounds the sample in the counting chamber; enhances discrimination against background and yields the highest E²/B values available in a multipurpose liquid scintillation counting system. It is specially designed for counting extremely low activity samples in disposable glass and plastic vials.
- **Super low level counting** is available for extremely low activity beta samples, increasing system sensitivity (E²/B) to a factory test minimum of 880 for ³H and 6,000 for ¹⁴C.
- **Multi-parameter linear MCA (Multichannel Analyzer)** with an effective resolution of 1/10 keV, offers an extended dynamic quench range and provides multi-parameter spectrum analysis to correct for luminescence, color quenching and background radiation.
- **SpectraBase counting and data management system** provides regionless counting and storing of complete spectra for all samples and standards. Features include automatic recall of spectra stored in the quench library for region-independent quench correction and post-processing of sample data.
- **A cassette-loaded bi-directional sample conveyor mechanism** is standard with a sample capacity of either 408 standard 20 mL vials, or 720 small 4 or 7 mL vials with Varisette™ sample changer for inter mixing large and small vial cassettes with out requiring special adapters.
- **Positive sample identification** provides protocol number, cassette number, sample number, and user-selectable printout and data file storage of the counting time and date for each sample.



Tri-Carb® 3170 TR/SL



www.perkinelmer.com/lifesciences

- **Quick-Count sample loading** for 60 independent protocols (with unlimited assays) provides unrestricted access to sample changer and protocol selection plugs. Sample batches are processed by simply activating the Quick-Count protocol plugs, minimizing any user programming.
- **Enhanced Replay sample post processing** provides complete recall and post processing of historical count data to eliminate sample recounting. It enables changes to count conditions and reports and executes user application software for optimization of data analysis. (Direct DPM results are not available for replay.)
- **Enhanced IPA™ (Instrument Performance Assessment) database** monitors efficiencies, backgrounds, E²/B and Chi-square values for ³H and ¹⁴C over the life of the instrument. IPA flags impending problems and provides both running mean and fixed baseline charts and associated tables for retrospective quality control and pro-active system maintenance. Baseline acquisition is programmable for increased flexibility. IPA data may also be transmitted to an external computer via RS-232 or saved to disk for archiving instrument performance. The most recent IPA time and date stamped data are available on demand for reporting purposes. Each IPA printout includes instrument model, serial number, software version number and calibration standard information.
- **SIS (Spectral Index of Sample) quench indicating parameter** determines counting efficiency by analysis of sample spectrum.
- **¹³³Ba low energy external standard source and tSIE (transformed Spectral Index of External standard) calculations** eliminate the effects of vial glow, plastic wall and cocktail changes on the DPM results. The use of integral spectrum counts eliminates the need for repeat counting of the external standard and negates the effect of isotope half-life on quench monitoring accuracy and precision. The ¹³³Ba external standard is centered under the sample vial which eliminates the affects of volume variations and assures reproducible quench monitoring for the life of the instrument.
- **AEC (Automatic Efficiency Control)** corrects for differential quenching effects in multi-label samples. The low energy spectrum of the external standard ensures accurate tracking of ³H, ¹⁴C and other low energy sample spectra over a very wide quench range.
- **Quench standard spectra** are factory-installed for single and dual label counting of ³H and ¹⁴C samples. Quench standard spectra are stored for both Ultima Gold™ liquid scintillation cocktail samples and toluene or xylene based samples.
- **Dynamic color-corrected single and dual label DPM** based on tSIE/AEC includes DPM based on SIS, constant quench DPM, and full spectrum DPM based on spectrum unfolding.
- **Triple-label DPM (Disintegrations Per Minute)** is based on tSIE/AEC for accurate spill correction.
- **Chemiluminescence correction with response normalization** corrects for luminescence interferences to speed up sample counting. Response normalization of the correction circuits eliminates the effect of component drift on the results.
- **Enhanced Direct DPM** technique determines the DPM of any single label pure beta or beta/gamma radionuclide in any cocktail without the use of quench standards.
- **Live SpectraView™ automatic spectrum display** aids in optimizing counting conditions, and helps evaluate complex sample situations. It allows setting temporary regions on the spectrum screen, and enables the operator to monitor the effect of AEC while the sample is counting.
- **Automatic spectrum plot** (on demand) per sample allows spectral documentation of samples.
- **Sample worklist and bar code reading** enables entry, editing and review of worklists for each assay. This allows sample identification by user-specific codes for sample printouts and data files. The use of a wedge-type bar code reader (not provided) to enter sample identification is accommodated.
- **Password protection** prevents unwanted changes to saved assays.
- **Printed header contains instrument serial number, user ID, and drive and path of stored data.** Each printed page or RTF (Rich Text Format) file report is numbered and dated for GLP compliance.
- **Assay-specific, user selectable, coincidence resolving time** enables optimized counting of any fast or slow, liquid or solid scintillator, for SPA (Scintillation Proximity Assays), solid scintillation filters and the newer liquid scintillation cocktails.
- **On-screen editing** of quench correction curves and recording of date last modified is present.
- **Electrostatic controller** automatically helps neutralize static electricity on all types of vials.
- **Automatic processing of count data to final results** provides automatic, protocol specific data processing for all user applications, including commercial or user-generated software. No exporting of data to off-board storage devices or computers is required.
- **Spectral unfolding**, separates and displays in color the individual radionuclide spectra of dual label samples.
- **3-D (three-dimensional) spectral mapping** displays in color the quench standard spectra together with the spectrum of the unknown for single label DPM counting.
- **Nuclide identification** is available for most common radionuclides.
- **Automatic region finder** with spectrum-search capability determines optimum region settings.
- **Auto-optimization of counting regions** maximizes counting sensitivity for ultra low activity samples.
- **Half-life correction to any date and time** is available for up to three radionuclides.
- **Activity reporting** is provided in Becquerels, microCuries, or picoCuries.
- **Custom output formatting to printer, RS-232 and disk storage** for each protocol provides almost unlimited flexibility in data reporting. Electronic data can be saved to disk in comma delimited ASCII, RTF, Lotus or Excel formats.
- **User-definable calculations** are available for custom data reporting.
- **Background subtraction** can be nominated via sample, entered value, or stored IPA background spectrum.
- **Programmable single photon counting** enables luminescence assay counting with optimized signal-to-background ratios. It overcomes problems associated with excessive luminescence.
- **Preset time (up to 9,999.99 minutes) and preset error coincidence termination** optimizes counting accuracy in the three counting regions.
- **Percent of standard calculations** is present for single, dual and triple label samples.
- **User adjustable assay-specific sample precount delay** permits dark adaptation of samples before counting.
- **Group PrioStat™ interrupt mode** gives priority counting status to a batch of samples counted according to any stored protocol conditions. It automatically restores the interrupted protocol upon completion and stores PrioStat data for immediate viewing. Data is printed at protocol termination.
- **Sample PrioStat™ interrupt mode** allows special function priority counting of individual samples, with manual control over counting conditions.
- **Anti-jam recovery** of the sample changer mechanism protects samples, vials and the counting system from damage if obstructions occur.
- **Automatic power-fail recovery** restarts counting when power is restored and the instrument has reinitialized itself.
- **Computer-aided diagnostics** are used to verify all system functions.
- **Heterogeneity monitor** determines sample quality and flags non-homogeneous sample results.
- **Decay computations** automatically calculate decay corrected DPM values for commonly used radionuclide standards.
- **Date and time clock** provides real time display and time-stamped printouts; battery supported.
- **Maneuverable CRT support arm** provides optimal positioning of the CRT monitor and conserves bench space.
- **Ink jet color printer** is standard.
- **SpectraWorks™ spectrum analysis software for the Windows® operating system** analyzes beta, alpha, and gamma spectra and provides simultaneous display for up to four spectra in stacked or overlaid mode. It features zooming to any part of the spectrum; six regions of interest; display of counts or CPM and linear or log spectra; provides automatic and manual scaling; calculates E²/B, MDA, peak resolution; allows adding and subtracting of spectra and multiplication and division by constants.
- **Temperature-controlled refrigeration** establishes and maintains optimum counting conditions for a wide variety of sample types.

Options

- **Ethernet adapter kit** includes an Ethernet card, internal cabling and connector panel. It provides simple connection to networks. The customer must supply and install network software driver.
- **Alpha/beta discrimination** using automatic TR-PDA (Time-Resolved Pulse Decay Analysis) separates alpha and beta radionuclides including automatic and manual optimization of minimum spillover settings applied to each protocol. It includes automatic determination, display, plotting and storage of alpha-in-beta and beta-in-alpha spillover curves. It also enables further optimization or fine-tuning of minimum spillover settings for each protocol and library storage of PDD (Pulse Decay Discriminator) setting for recall.