

UVP Translinker

The UVP Translinker combines the shortwave UV CL-1000 Crosslinker and the M-20 Transilluminator. The transilluminator is built onto the crosslinker cabinet.

- On/Off controls for the transilluminator are located on the crosslinker panel
- The transilluminator uses midrange 302 nm UV and a 20 x 20 cm filter



Order Number		Description
115 V	230 V	UVP Crosslinkers, UVP Translinkers and Replacement Tubes
95-0174-01	95-0174-02	CL-1000 Crosslinker
95-0228-01	95-0228-02	CL-1000L Crosslinker
95-0230-01	95-0230-02	CL-1000M Crosslinker
95-0339-01	95-0339-02	CX-2000 Crosslinker
95-0300-01	95-0300-02	TL-2000 Translinker
	34-0007-01	Replacement Tube (CL-1000, CX-2000, TL-2000), 8-watt, 254 nm UV
	34-0006-01	Replacement Tube (CL-1000L), 8-watt, 365 nm UV
	34-0042-01	Replacement Tube (CL-1000M, TL-2000), 8-watt, 302 nm UV

UVP Incubator

This UV sterilizing incubator provides precise temperature control and uniformity throughout the chamber for incubation of biological assays, fungal, bacterial cultures, eggs, and other samples.



- Microprocessor-controlled incubation temperature
- Heating element and fan generate uniform temperature throughout the chamber
- Decontaminate the chamber between growth experiments with the overhead shortwave 254 nm UV sterilization
- Two ventilated shelves included; remove the shelves to stack up to 150 (10 cm) plates
- Temperature Ambient to 68°C, Accuracy: +0.5°C
Uniformity: +0.1°C at 37°C, Humidity: ~80%

Order Number		Description
115 V	230 V	UVP Incubator and Replacement Tube
95-0358-01	95-0358-02	UV Benchtop Incubator (SI-950)
	34-0007-01	Replacement Tube, 8-watt, 254 nm UV



Wolflabs

Wolf Laboratories Limited

www.wolflabs.co.uk

Tel: 01759 301142

Fax: 01759 301143

sales@wolflabs.co.uk



Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

