

Orbital and Linear Digital Shaker

User Manual



SK-L330-Pro Linear Digital shaker (7.5Kg)

SK-O330-Pro Orbital Digital shaker (7.5Kg)

SK-L180-Pro Linear Digital shaker (2.5Kg)

SK-O180-Pro Orbital Digital shaker (2.5Kg)

Please read the User Manual carefully before use, and follow all operating and safety instructions!

Technical specifications and outline are subject to change without prior notice.

www.dlabsci.com



Contents

| | |
|------------------------------------|----|
| Preface | 1 |
| Service | 1 |
| Warranty | 1 |
| 1. Safety Instructions | 2 |
| 3. Inspection | 3 |
| 3.1 Receiving Inspection | 3 |
| 3.2 Listing of Items | 3 |
| 4. Trial Run | 4 |
| 5. Control and Set | 5 |
| 5.1 Control | 5 |
| 5.2 Set | 6 |
| 5.2.2 Set speed | 6 |
| 6. Operating Modes | 7 |
| 6.1 Mode A | 7 |
| 6.2 Mode B | 8 |
| 6.3 Switching the mode | 9 |
| 7. Supported Load | 9 |
| 8. Faults | 9 |
| 9. Interface | 9 |
| 10. Maintenance and Cleaning | 10 |

| | |
|--|----|
| 11. Associated Standards and Regulations | 10 |
| 12. Specifications | 11 |
| 13. Products and Accessories | 12 |

Preface

Welcome to the “Orbital and Linear Digital Shaker User Manual”. Users should read this manual carefully, follow the instructions and procedures, and beware of all the cautions when using this instrument.

Service

When help needed, you can always contact manufacturer for technical support in the following ways:

DLAB Scientific Instrument Inc.

2311 E. Locust Court, Ontario, CA 91761 United States.

Office: +1-747-230-5179

Fax: +1-909-230-5275

Sales contact: info@dlabsci.com

Service contact: service@dlabsci.com

www.dlabsci.com



Please provide the customer care representative with the following information:

- Serial Number (on the rear panel)
- Certification
- Description of problem (i.e., hardware or software)
- Methods and procedures adopted to resolve the problems
- Your contact information

Warranty

This instrument is warranted to be free from defects in materials and workmanship under normal use and service, for a period of 24 months from the date of invoice. The warranty is extended only to the original purchaser. It shall not apply to any product or parts which have been damaged on account of improper installation, improper connections, misuse, accident or abnormal conditions of opera local dealer. You may also send the instrument direct to our works, enclosing the invoice copy and by giving reasons for the claim. You would be solely liable for freight costs.

1. Safety Instructions

| | |
|--|---|
|  | <p>Warning!</p> <ul style="list-style-type: none"> • Read the operating instructions completely before use and follow the safety instructions. • Ensure that only trained staff work with the instrument. |
|  | <p>Protective ground contact !</p> <ul style="list-style-type: none"> • Make sure the socket must be earthed (protective ground contact) before use. |

- Wear the personal protective equipment in accordance with the hazard category of the media to be processed.

-Splashing of liquids.

-vibration force led to broken glass container.

-The body, hair, clothes, jewelry are inhaled.

- Use the instrument following the safety instructions and profession safety rules to avoid any accident.
- Users are not allowed to touch moving parts of the instrument to avoid any accidents. Pay attention to protect hands or fingers from rolling when moving the instrument.
- Set up the instrument in a spacious area on a stable, clean, non-slip, dry and fireproof surface. Do not operate the instrument in explosive atmospheres, with

hazardous substances or under water.

- Observe the vessel on the shaking table for medium splashing out of liquid when setting the motor speed. Reduce the motor speed if the instrument is not running smoothly.
- Accessories must be securely attached to the instrument so that it cannot come off by itself and vessels must be firmly placed on the shaking table to avoid any injury.
- Working with combustible media or flammable materials may result in additional hazards. Only process media that will not react dangerously to the extra energy produced through processing should be used on the instrument.
- Safe operation is only guaranteed with the accessories described in the “Accessories” chapter and by following the operating instructions in full. Always disconnect the plug before fitting accessories. Check the instrument and accessories beforehand for damage each time you use them. Do not use damaged components.
- The instrument may only be opened by trained technicians. And always disconnect the plug

beforehand.

- The operating voltage required as stated on the instrument and voltage in the supply network must match.
- Do not cover the instrument. Protect the instrument and accessories from bumps and impacts.
- Keep away from high magnetic field.



Note:

If there is any apparent damage to the system, please do not plug it into the power line.

2. Proper Use

The instrument is designed for mixing or shaking liquids in schools, laboratories or factories. SK-330/180-Pro is suitable for usage in various attachments for mixing liquids in bottles, flasks, test tubes for a maximum supported weight of 7.5 kg (330) and 2.5Kg (180). This instrument is not suitable for use in residential areas or other areas that may cause danger to the user or instrument as mentioned in chapter 1.

3. Inspection

3.1 Receiving Inspection

Unpack the instrument carefully and check for any damages which may have arisen during transport. If it happens, please contact manufacturer for technical support.

3.2 Listing of Items

The packing includes the following items:

| Items | Qty |
|-------------|-----|
| Main unit | 1 |
| Power cable | 1 |
| User manual | 1 |

Table 1

4. Trial Run

Do trial operation as follows:

- Make sure the required operating voltage and power supply voltage match.
- Ensure the socket is securely earthed.
- Plug in the power cable, Power ON the instrument.
- LCD displays the safe rotary speed limit.
- LCD displays operating mode.
- Turn the speed knob to set the rated rotary speed
- Press the speed knob. And the shaking function is switched ON.
- Press the motor regulation knob again. and the shaking function is switched OFF.

If these operations above are normal, the instrument is ready to operate following the “operation” chapter. If these operations are not normal, the instrument may be in security protection state or be damaged.

If the mounting surface is not even, you can improve the stability of the instrument with the adjustable feet. To do this, turn the appropriate foot downward with the wrench

until the instrument is standing securely on the surface.




Figure 1

5. Control and Set

5.1 Control




Figure 2

| Items | Descriptions |
|------------|---|
| Speed knob | Set rated rotary speed, “rotation” is switched ON or OFF via pressing the knob. |
| Time knob | Set working time, “rotation” is switched ON or OFF via pressing the knob |
| LCD | Display the status of instrument and any |

| | |
|--------------|-------------------|
| | set |
| Mains switch | Switch ON or OFF. |

Table 2

- Put the instrument on stable and safe place and plug in the mains power.
- Turn ON the mains switch on the left panel.
- The instrument starts self-test.
- The instrument shows rated speed and time after initialization.
- Turn the speed knob on the left side to set the rated speed.
- Press the speed knob and characters on the left LCD do not flash any more, and the shaking function is switched ON.
- Press the speed knob again and characters on the left LCD begin to flash, and the shaking function is switched OFF.
- Turn the time knob on the left side to set the rated time
- Press the time knob and characters on the LCD do not flash any more, and the timing function is switched

ON.

- Press the time knob again and characters on the right LCD begin to flash, and the timing function is switched OFF.

5.2 Set

5.2.1 Set time

The rated shaking time can be set by turning the time knob on the right. A distinction is made for the shaking time setting between timed mode and continuous operation. If continuous mode is selected, the shake can continue its shaking function for any amount of time with the previously set speed. If timed mode is selected, the shaker can run in the set time. After the instrument is powered OFF and restarted, the set time is erased to zero blanking and it switches into offline operating mode.

If a target time (max. 19h 59min) and speed are set, users can activate the instrument with pressing the time knob, and the time begin to read, If:

- A, pressing the time knob, the speed and the time will stop. Pressing the time knob again, restart the shake function, the time will use the pre-set time.

- B, pressing the speed knob, the speed and the time will stop; Pressing the time knob again, restart the shake function, the time not (There are flashing on the right of the LCD), pressing the time knob again restart the time function. The time will use the pre-set time.



Note:

The current set time can be varied at any time.

5.2.2 Set speed

The rated speed and upper speed limit can be set by turning the speed knob (As the Figure 6). The continuous mode can be switched ON with pressing the left knob without the time setting and switched OFF with pressing the knob again.

Adjust the motor speed knob slowly in order to keep the instrument running smoothly.



Note:

The current set speed can be varied at any time.

6. Operating Modes

6.1 Mode A

Operating mode A is the default mode set in factory. After turning ON the instrument with the mains switch, the shaking and timing functions are turned OFF. The LCD displays the set time and speed values. They are adopted or varied when the corresponding functions are turned ON.

After turning ON the instrument, the following process will be shown on the display.

- The left LCD displays “SAF” and the right displays the upper speed limit(rpm) which can be set by pressing the speed knob and turning at the same time.



Figure 3

- The LCD displays run modes “StA”、 “A” or “B” for about 2 seconds.



Figure 4

- The set or stored speed value is shown in the left display, and the set or stored time value is shown in the right display. Turn the speed knob and time knob can set the speed and time.




Figure 5

- After pressing the speed knob or time knob, the instrument begins to run at the set speed. The actual speed and time remaining are shown in the LCD. If the remaining has reached zero, the shaker stops its motion. Pressing the speed knob alone, the instrument will work at continuous mode (The time function is OFF)



Figure 6

6.2 Mode B

When the instrument is turned ON and values for speed and time are erased to zero, users can set. The upper speed limit that has been set is taken over from operating mode A and cannot be modified. After the power has been turned OFF, the instrument will no longer automatically start up in operating Mode B by itself.

After turning ON the instrument and selected run mode, the following process will be shown on the display.

- The left LCD displays “SAF” and the right displays the upper speed limit(rpm) which cannot be changed.




Figure 7

- The LCD displays run modes “StA” for 2 seconds.




Figure 8

- The set speed value is shown in the left display, and

the set time value is shown in the right display. Then set the motor rotary speed and time.




Figure 9




Figure 10

- After pressing the speed knob or time knob, the instrument begins to run at the set speed. The actual speed and time remaining are shown in the LCD. If the remaining has reached zero, the shaker stops its motion.




Figure 11



Note:

The set values can be varied during shaking. Shaking can be stopped by pressing the left or right knob.

6.3 Switching the mode

Switch the operating mode as follows:

- Turn OFF the instrument with the mains switch.
- Hold down both knobs and turn ON the instrument with the mains switch. After 5 second, you can let go of the knobs.
- Switch between operating Modes A and B in order.

7. Supported Load

In order to ensure safe, the shaker must only be operated within the supported weight.

Make sure the holding surface of the shaker is always clean and level. Ensure the individual shaking vessels must be placed evenly and fastened securely in the middle of the shaking table no matter one or multiple shaking vessels are placed.

8. Faults

- When switched ON, the instrument doesn't work
 - Check whether the power cable fitted well
 - Check whether the fuse is broken

- The speed cannot reach set value
 - Check whether it is overloaded
- The motor does not start via pressing the speed knob and time knob
 - Check whether the time is set to zero

9. Interface

The Orbital and Linear Digital Shaker, special adapter and cable are used to link to external device. and the standard 9-pin interface to connect with the PC.

- The cable between the shaker and the computer used EIA Standard RS232C communication line, corresponding to the DIN 66020 interface.
- Transmission method: Asynchronous signal transmission.
- Mode of transmission: Fully Duplex.
- 1 start bit; 8 character bits; 1 stop bit.
- Transmission speed: 9600 bit/s



Note:

Forbid to plug in and off the power cable when instrument works.

10. Maintenance and Cleaning

Use and maintain the instrument in right way to keep it in good operating state in order to lengthen its life time and avoid the damage to the instrument. Only use the cleaning agents which have been approved by manufacturer to clean the instrument.

Only use cleanser that we advised as below:

| | |
|------------------------|--|
| Dyes | Isopropyl alcohol |
| Construction materials | Water containing tenside / isopropyl alcohol |
| Cosmetics | Water containing tenside /isopropyl alcohol |
| Foodstuffs | Water containing tenside |
| Fuels | Water containing tenside |

For materials which are not listed, please request information from manufacturer or your supplier. Before using another than the recommended method for cleaning or decontamination, the user must ascertain that this method does not destroy the instrument. Wear the proper protective gloves during cleaning of the instrument.



Note:

- Electrical instrument may not be placed in the cleansing agent for the purpose of cleaning.
- The instrument must be cleaned and put it into the initial packaging carton before sending to service for repair, avoiding the contamination of hazardous.
- Switch OFF the instrument and put it in the dry, clean, stable place at room temperature in long-term disuse.

11. Associated Standards and Regulations

Construction in accordance with the following safety standards

EN 61010-1

UL 3101-1

CAN/CSA C22.2(1010-1)

EN 61010-2-10

Construction in accordance with the following EMC standards

EN 61326-1

12. Specifications

| Items | Specifications |
|--|--|
| Voltage [VAC] | 100~240 |
| Frequency [Hz] | 50/60 |
| Power [W] | 30 |
| Shaking movement | Orbital/ linear |
| Orbital diameter [mm] | 10 (SK-O(L)330-Pro) 4.5 (SK-O(L)180-Pro) |
| Max. shaking weight (with attachment) [kg] | 7.5 (SK-O(L)330-Pro) 2.5 (SK-O(L)180-Pro) |
| Motor type | External rotor brushless motor |
| Motor rating input [W] | 28 |
| Motor rating output [W] | 15 |
| Permissible ON time [%] | 100 |
| Speed range [rpm] | 100-350(SK-L330-Pro) 100-500(SK-O330-Pro) 100-350(SK-L180-Pro) 100-800(SK-O180-Pro) |
| Speed display | LCD |

| Items | Specifications |
|---------------------------------------|--|
| Timer | Yes |
| Timer display | LCD |
| Time setting range [min] | 1-1199 |
| Run type | Time / Continuous operation |
| Dimensions [D×W×H mm] | 420×370×100(SK-O(L)330-Pro) 340×300×100(SK-O(L)180-Pro) |
| Weight [kg] | 13.5(SK-O(L)330-Pro) 8.1(SK-O(L)180-Pr |
| Permissible ambient temperature [°C] | 5-40 |
| Permissible relative humidity | 80% |
| Protection class acc. To DIN EN60529 | IP21 |
| RS232 interface | Yes |

Table 3

13. Products and Accessories

| Cat. No. | Descriptions |
|--------------|---|
| 831010017777 | SK-L180-Pro Linear Digital Shaker (2.5Kg),without attachment, Cn plug, 110/220/50Hz/60Hz |
| 831010037777 | SK-L180-Pro Linear Digital Shake (2.5Kg), without attachment, Euro plug,110/220/50Hz/60Hz |
| 831010047777 | SK-L180-Pro Linear Digital Shaker (2.5Kg), without attachment, UK plug, 110/220/50Hz/60Hz |
| 831010057777 | SK-O180-Pro Orbital Digital Shaker (2.5Kg), without attachment, Cn plug, 110V/220V/50Hz/60Hz |
| 832010017777 | SK-O180-Pro Orbital Digital Shaker (2.5Kg), without attachment, Euro plug, 110V/220V/50Hz/60Hz |
| 832010037777 | SK-O180-Pro Orbital Digital Shaker (2.5Kg) , without attachment, Euro plug, 110V/220V/50Hz/60Hz |
| 832010047777 | SK-O180-Pro Orbital Digital Shaker |

| | |
|--------------|---|
| | (2.5Kg) , without attachment, UK plug, 110V/220V/50Hz/60Hz |
| 832010057777 | SK-O180-Pro Orbital Digital Shaker (2.5Kg) , without attachment, USA plug, 110V/220V/50Hz/60H |
| 831010027777 | SK-L330-Pro Linear Digital Shaker (7.5 Kg), No Attachment, Cn plug, 110V/220V/50Hz/60Hz |
| 831010067777 | SK-L330-Pro Linear Digital Shaker (7.5 Kg) , without attachment, Euro plug, 110V/220V/50Hz/60Hz |
| 831010077777 | SK-L330-Pro Linear Digital Shaker (7.5 Kg) , without attachment , UK plug, 110V/220V/50Hz/60Hz |
| 831010087777 | SK-L330-Pro Linear Digital Shaker (7.5 Kg) , without attachment, USA plug, 110V/220V/50Hz/60Hz |
| 832010027777 | SK-O330-Pro Orbital Digital Shaker (7.5 Kg) , without attachment, Cn plug, 110V/220V/50Hz/60Hz |
| 832010067777 | SK-O330-Pro Orbital Digital Shaker (7.5 Kg) , without attachment, |

| | |
|--------------|---|
| | Euro plug, 110V/220V/50Hz/60Hz |
| 832010077777 | SK-O330-Pro Orbital Digital Shaker (7.5 Kg) , without attachment, UK plug, 110V/220V/50Hz/60Hz |
| 832010087777 | SK-O330-Pro Orbital Digital Shaker (7.5 Kg) , without attachment, USA plug, 110V/220V/50Hz/60Hz |

Accessories

| Accessories | Cat. No. & P/N | Descriptions |
|--|---------------------|--|
|  | 18900025 SK180.1 | Universal attachment |
|  | 18900026 SK180.2 | Fixing clip attachment |
|  | 18900038 SK180.3 | Lengthways roller Attachment |
|  | 18900039 SK180.4 | Dish attachment (with slip-resistant foil) |

| | | |
|---|------------------------------------|--|
|  | 18900027 SK330.1 | Universal attachment |
|  | 18900028 SK330.2 | Fixing clip attachment |
|  | 18900040 SK330.3 | Lengthways roller Attachment |
|  | 18900041 SK330.4 | Dish attachment (with slip-resistant foil) |
|  | 18900029 SK180.2.1 SK330.2.1 | Fixing clip for flask volume 25 ml |
|  | 18900030 SK180.2.2 SK330.2.2 | Fixing clip for flask volume 50 ml |
|  | 18900031 SK180.2.3 SK330.2.3 | Fixing clip for flask volume 100 ml |

| Accessories | Cat. No. P/N | Descriptions |
|--|---|---|
|  | 18900032 SK180.2.4 SK330.2.4 | Fixing clip for flask volume 200/250 ml |
|  | 18900033 SK180.2.5 SK330.2.5 | Fixing clip for flask volume 500 ml |
|  | 18900036 | Bar for SK330.1 Universal attachment, black |
|  | 18900037 | Bar for SK180.1 Universal attachment, black |
| | 18100319 RS232 9-Pin interface line | Suitable to ShakerPC Software |

Note:

The software ShakerPC can be downloaded free on our website www.dragon-lab.com



DLAB Scientific Instrument Inc.

2311 E. Locust Court, Ontario, CA 91761 United States.

Office: +1-747-230-5179

Fax: +1-909-230-5275

Sales contact: info@dlabsci.com

Service contact: service@dlabsci.com

www.dlabsci.com