Key Features

- Conductivity, resistivity, TDS and salinity modes
- Special pure water mode for ultra low conductivity measurements
- Storage of 500 results with built in data logger
- Infra-red data link (IrDA) for connection to printer and RS232 link to connect to printer or PC via DataWay software



Part code: 452 001



Bench Conductivity/TDS Meter

The 4520 is a high specification laboratory conductivity /resistivity/ TDS/salinity and temperature meter that offers additional accuracy with 1, 2 or 3 point conductivity calibration across the wide measurement range of 0 to 19.99S. The dedicated 'pure water' mode ensures the optimum accuracy for this difficult application.

Technical Specification

| Conductivity | |
|-------------------------------|------------------------------------|
| Range | 6 auto-selected from 0 to 19.99S* |
| Resolution | 0.01µS -0.01S |
| Accuracy | ±0.5% ±2 digits |
| Automatic std. recognition | 10µՏ, 84µՏ, 1413µՏ, 12.88mՏ |
| Cell constant | 0.010 to 19.999 |
| TDS | |
| Range | 6 auto selected ranges 0-1999g/l* |
| Resolution | 0.01mg/l -1g/l |
| Accuracy | ±0.5% ±2 digits |
| EC ratio | 0.50 to 0.80 |
| Resistivity | |
| Range | 0-20Mohm |
| Resolution | 0.01Mohm |
| Accuracy | ±0.5% ±2 digits |
| Salinity | |
| Range | 0-99.9g/l |
| Resolution | 0.1g/l |
| Accuracy | 0-35 ±1g/l, 35-99.9 ±3g/l |
| Temperature | |
| Range | -10 to 105°C |
| Resolution | 0.1°C |
| Accuracy | ±0.5°C |
| ATC and manual | 0 to 100°C |
| Temperature coefficient | 0.00% to 4.00%/°C |
| Reference temperature | 18, 20 or 25°C |
| GLP | Calibration reminder interval, |
| | alarm outputs and security code |
| | protected data |
| Outputs | Analogue, RS232 and IrDA interface |
| Connector | 7-pin DIN |
| Power | 9V AC ±10% @ 50/60Hz** |
| Size (l x w x d), mm | 210 x 250 x 55 |
| Weight, g | 850 |
| *With K=10 cell | |
| ** Voltage variants available | e see page 94 |
| | |

Ordering Information

| Part Code | Description |
|-----------|--|
| 452 001 | 4520 conductivity meter supplied with glass |
| | conductivity probe with ATC (K=1) (027 013), |
| | electrode holder (903 300) and UK power supply |
| | (021 030) |



Wolf Laboratories Limited

www.wolflabs.co.uk

Tel: 01759 301142

Fax:01759 301143

sales@wolflabs.co.uk







Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.





