

Ziegelei 1 D-72336 Balingen E-mail: info@kern-sohn.com Tel: +49-[0]7433- 9933-0 Fax: +49-[0]7433-9933-149 Internet: www.kern-sohn.com

# **User instructions** Stereo microscope set





# **KERN OZM-98**

Version 1.0 01/2016 User instructions Stereo microscope set

# Table of contents

1	Before use	
1.1 1.2 1.3 1.4	General notes Notes on the electrical system Storage Maintenance and cleaning	
2	Nomenclature	6
3	Basic data	8
4	Assembly	9
5	Operation and functionality	15
5.1	Getting started	
5.2	Adjust the interpupillary distance	15
5.3	Adjusting the magnification	15
5.4	Dioptre adjustment and focussing	
5.5	Adjusting the stand	
<b>J.O</b> 5.7	Brightness adjustment	
5.8	Fitting and adjusting a camera	
5.9	Using additional accessories	
6	Optical data	23
7	Trouble shooting	23
8	Service	24
9	Disposal	24
10	Further information	24

#### 1 Before use

#### 1.1 General notes

You must open the packaging carefully, to make sure that none of the accessories in the packaging fall on the floor and get broken.

In general, microscopes should always be handled carefully because they are sensitive precision instruments. When using or transporting the microscope it is particularly important to avoid abrupt movements, as this may damage the optical components.

You should also avoid getting dirt or finger prints on the lens surface, because in most cases this will reduce image clarity.

To maintain the performance of the microscope, it must never be disassembled. So components such as lenses and other optical elements should be left as they were before use. Also the electrical parts of the device must not be tampered with, as in this area there is an additional risk of triggering an electric shock.

#### **1.2** Notes on the electrical system

Before connecting to a mains power supply, you must make sure that you are using the correct input voltage. The information to select the correct mains cable is located on each device, on the rear of the product directly above the connection socket. You must comply with this information. If you do not comply with these specifications, then fires or other damage to the device could occur.

The main switch must also be switched off before the mains cable is connected. In this way you will avoid triggering an electric shock.

If you are using an extension cable, then the mains cable you use must be earthed.

When carrying out any procedures whereby you come into contact with the electrical system of the device, such as, for example, changing the bulb or fuse, only carry out these procedures when the power is disconnected.

#### 1.3 Storage

You should ensure that the device is not exposed to direct sunlight, temperatures which are too high or too low, vibrations, dust or a high level of humidity.

The ideal temperature range is between 0 and 40°C and a relative humidity of 85% should not be exceeded.

The device should always be located on a rigid, smooth, horizontal surface.

For devices with pillar stands, the microscope holder must not be rotated back too far. If you do this, there is a risk that the microscope could tip over.

When the microscope is not being used, you should fit the objective cap and cover the microscope with the enclosed dust protective cover.

If the eyepieces are being stored separately, the protective caps must be fitted to the tube connectors. In most cases, if dust and dirt gets inside the optical unit of a microscope this can cause irreversible errors or damage.

The best way to store accessories which consist of optical elements, such as, for example, eyepieces and objectives, is in a dry box with desiccant.

#### 1.4 Maintenance and cleaning

In any event, the device must be kept clean and dusted regularly.

If any moisture should be occur, before you wipe down the device you must ensure that the mains power is switched off (if lighting unit is fitted).

When glass components become dirty, the best way to clean them is to wipe them gently with a lint-free cloth.

To wipe oil stains or finger prints off the lens surface, moisten the lint free cloth with a mixture of ether and alcohol (70 / 30 ratio) and use this to clean the lens.

You must be careful when handling ether and alcohol, as these are highly flammable substances. You must therefore keep it away from naked flames and electrical devices which can be switched on and off, and only use it in well-ventilated rooms.

However organic solutions of this type should not be used to clean other components of the device. This could lead to damage to the paint finish. To do this, it is sufficient to use a neutral cleaning product.

You could also use the following cleaning products to clean the optical components:

- Special cleaner for optical lenses
- Special optical cleaning cloths
- Bellows
- Brush

When handled correctly and checked regularly, the microscope should give many years of efficient service.

Should repairs still be necessary, please contact your KERN dealer or our Technical Department.

#### 2 Nomenclature

Using the OZM 983 as an example (Trinocular)





### 3 Basic data

Optical system	Greenough
Dimmable lighting	Yes
Magnification ratio	6,4:1
Tube	angled at 45°
Interpupillary distance	52 – 76 mm
Dioptre adjustment	On both sides
Eyepiece	HSWF 10x / Ø 23mm
Field of view (mm)	Ø 32,8 – 5,1

## Standard configuration

Model	Microscope head		Stand	Holder	Illumination
KERN	Tube	Objective Zoom			mm
OZM 982	Binocular (OZM 546)	0,7×-4,5×	Spring loaded arm with clamp (OZB-A6302)	With double handle	4,5 W LED ring light (OBB-A6102)
OZM 983	Trinocular (OZM 547)	0,7×-4,5×	Spring loaded arm with clamp (OZB-A6302)	With double handle	4,5 W LED ring light (OBB-A6102)

# 4 Assembly

NO.	Article	Description	Quantity	
A		Mounting clamp - a	1	
В		Mounting clamp - b	1	
D		M8≉90	1	
Е		Microscope stand - lower part	1	
F	nici Ph	Micorscope stand - upper part	1	
G	(Esa (O)	guide handle	1	
Н	2	M4*10	3	
I		Microscope holder	1	
J		Microscope head	1	
K		Eyepieces	2	
L		Eye cups	2	
М		Мб	1	
N		М4	1	
0		МЗ	1	

OZM-98-BA-e-1610

















In order to ensure the incident illumination, the LED ring light needs to be attached to the bottom of the objective housing of the microscope head. The three fastening screws on the holder of the ring light are used for the mounting.

Therefor it is important to pay attention to the groove around the objective housing, which is intended to be the socket for the screws.

#### Additional optional attachments:

• You can fit a C-mount adapter to the appropriate connection point on the top of the microscope head for trinocular devices (OZM 983). This enables you to fit and use digital cameras (remove protective cap first) *(see section 5.8)*.

### 5 Operation and functionality

#### 5.1 Getting started

After assembly, if the microscope is ready for use, then for devices with lighting you must first establish a **power connection** using the cable supplied. *Please see section 5.7 for more details on adjusting the lighting.* 

If the device has a camera connection (OZM 983), then for **binocular** use, you must let the **trinocular toggle rod** on the back of the microscope head stay **slid-in**. *For trinocular use (camera connection) see section 5.8.* 

Do not forget to remove the cap from the bottom of the objective, so that you will then be able to see a reflection of the object being observed in the eyepiece.

All important functions which relate to the use of the devices in this document are described in the following sections.

#### 5.2 Adjust the interpupillary distance

Different users have different interpupillary distances. So each time a different person uses the microscope, the gap between the two eyepieces must be re-adjusted.

While you are looking through the eyepieces, use one hand to hold the righthand or lefthand prism housing firmly.

By rotating outwards or inwards, you can either increase or reduce the interpupillary distance.

As soon as the lefthand and righthand visual fields exactly overlap each other, this is the correct interpupillary distance.

#### 5.3 Adjusting the magnification

As the KERN OZM-9 series are stereo zoom microscopes, then you adjust the magnification using the two zoom adjustment wheels on the lefthand and righthand side of the microscope head.

Chapter 6 "Optical data" gives information on the possible overall magnification which the microscope can produce. It will also include the optional use of different eyepieces and auxiliary objectives.

#### 5.4 Dioptre adjustment and focussing

A special feature of stereo microscopes is that they are fitted with an optical unit which has a relatively high depth of field. In order to be able get the most benefit from this feature, each user must synchronise the focussing mechanisms for themselves.

The steps to do this are described in the following section.

- 1. Place the object to be observed on the working surface under the objective.
- 2. Put both dioptre adjustment rings into the starting position of 0.
- 3. Use the zoom control dials to set the smallest possible magnification.
- 4. Look through the right eyepiece with the right eye and bring the object into focus by using the focus control dials.
- 5. Now set the largest possible zoom factor.
- 6. Once again, still only looking through the right eyepiece, bring the object into focus
- 7. Then set the smallest possible zoom factor again.
- 8. If the object then does not appear to be in focus, adjust the focus on the dioptre adjustment ring of the right eyepiece.
- 9. In order to get the highest level of accuracy when adjusting the focus, you should repeat steps 5-8.
- 10. Afterwards set back to the smallest possible zoom factor.
- 11. Then look through the left eyepiece with the left eye and use the lefthand dioptre adjustment ring to also adjust the optimum focus of the object.
- 12. In this way, the object being observed will be in focus at any zoom setting.

#### 5.5 Adjusting the stand



All contained joints of the microscope stand can be application-orientated aligned and fixated (see upper picture).

The pneumatic spring ensures the height adjustment of the upper part of the microscope stand.

For varying the spring force, the adjustment screw on the middle joint of the stand must be used. Before performing this, the upper part needs to be kept in a completely depressed position (*see lower picture*).

#### 5.6 Using eye cups / High Eye Point eyepieces

The eye cups supplied with the microscope can basically be used at all times, as they screen out intrusive light, which is reflected from light sources from the environment onto the eyepiece, and the result is better image quality.

But primarily, if eyepieces with a high eye point (particularly suitable for those who wear glasses) are used, then it may also be useful for users who don't wear glasses, to fit the eye cups to the eyepieces.

These special eyepieces are also called High Eye Point eyepieces. They can be identified by the glasses symbol on the side. They are also marked in the item description by an additional "H" (example: HSWF  $10x \ 023 \ mm$ ).

When fitting the eye cups, make sure that the dioptre setting is not moved. We would therefore advise that you hold the dioptre adjustment ring on an eyepiece with one hand while you fit the eye cup with the other.

Before using the microscope, users who wear glasses must remove the eye cups, which you may find on High Eye Point eyepieces.

As the eye cups are made of rubber, you must be aware that when you are using them, they can become slightly dirty through grease residues. In order to maintain hygiene, we would therefore recommend that you clean the eye cups regularly (e.g. with a damp cloth).



Eye cups



High Eye Point eyepiece (identified by the glasses symbol)

#### 5.7 Brightness adjustment

There is only the usage of incident illumination available when working with KERN stereo microscope sets. Therefor a LED ring light (*see picture below*) is required, that can be mounted to the bottom of the microscope head (*see chapter 4 Assembly*).

Provided that the power connection is established, the illumination can be switched on by using the main switch, which is located at the housing rear of the illumination unit.

Furthermore it is possible to adjust the light intensity. Therefor you have to use the control knob on the side of the housing. Depending on which way you turn the knob, the light will become lighter or darker.



Control elements for the lighting unit

# 5.8 Fitting and adjusting a camera (OZM 983)



You can connect special microscope cameras and reflex cameras to trinocular devices in the OZM-9 series, so that you can digitally record images or sequences of objects being observed.

The connection for this is on the top side of the microscope head (see figure above).

When the **trinocular toggle rod is pulled out**, the beam path, which normally emits on the righthand eyepiece, is diverted in the tube, so that it can be used for the camera adapter connection.

This means that when the device is used in trinocular mode, you will see one image in the left eyepiece and one on the reflex camera or monitor of the microscope camera.

This means that the 3D effect is lost.

# To fit a <u>microscope camera</u> properly, you must use an adapter with a C-mount thread, which is put in the adapter connection once the cap is removed.

In total there are three focusable adapters to choose from *(see figure below)*. The difference between these adapters is that they have different integrated magnification (0.3x, 0.5x, 1.0x).

#### The camera and adapter are then united using the C-mount thread.

For special measuring applications in conjunction with a microscope camera, when using the 1.0x C-mount, there is the option of screwing in a second adapter with an integrated micrometre in front of the camera.



C-mount adapter

You also need an adapter to connect a <u>reflex camera</u>. But in contrast to the microscope camera this does not have a C-mount thread, but has a bayonet connection which fits directly on the camera objective.

There are also three different adapters to choose from. They all have a 1x magnification and differ in the fact that they come from different camera manufacturers. **Canon, Nikon and Olympus** cameras can be used for this.

The image which is shown on the camera connected to the device can often have a different level of focus compared with the image on the eyepiece, when **adapters** with integrated magnification (0,3x, 0,5x) are used. In order to be able to bring both images into focus, the focus can be adjusted by those adapters.

#### 5.9 Using additional accessories

Auxiliary objectives



In order to make the magnification series of one of the stereo zoom microscopes described here even more flexible, there is the option of using appropriate auxiliary objectives.

As you can see in Chapter 7 "Features", you can choose from five different achromatic corrected objectives (0.37x, 0.5x, 0.7x, 1.5x, 2.0x), depending on what is required.

You can fit these objectives by simply screwing them onto the existing standard objective with 1.0x magnification, which is already fixed to the microscope.

When you are doing this, you must avoid touching the objective lenses with your fingers or leaving any dust between the standard objective and the auxiliary objective.

# 6 Optical data

Eyepiece	Specifications – Objectives						
Magnification Standard		Auxiliary objectives					
		1,0x	0,37x	0,5x	0,7x	1,5x	2x
	Total magnification	7x - 45x	2,59x - 16,65x	3,5x - 22,5x	4,9x - 31,5x	10,5x - 67,5x	14x - 90x
HSWF TOX	Field of view mm	Ø 32,8 – 5,1	Ø 88,8 - 13,8	Ø 65,7 – 10,2	Ø 46,9-7,3	Ø 21,9 - 3,4	Ø 16,4 - 2,6
SW/E 15x	Total magnification	10,5x - 67,5x	3,89x - 25x	5,3x - 33,8x	7,4x - 47,2x	15,8x - 101,3x	21x - 135x
SWF 15X	Field of view mm	Ø 24,3-3,8	Ø 65,6 - 10,2	Ø 48,6-7,6	Ø 34,7 - 5,4	Ø 16,2 – 2,5	Ø 12,1-1,9
SW/E 20x	Total magnification	14x-90x	5,18x - 33,3x	7x - 45x	9,8x - 63x	21x - 135x	28x - 180x
3WF 20X	Field of view mm	Ø 20-3,1	Ø 54, 1 - 8,4	Ø 40-6,2	Ø 28,6-4,4	Ø 13,3-2,1	Ø 10-1,6
SWE 20x	Total magnification	21x-135x	7,77x - 50x	10,5x - 67,5x	14,7x - 94,5x	31,5x - 202,5x	42x - 270x
SWF SUX	Field of view mm	Ø 12,9-2	Ø 34,7 - 5,4	Ø 25,7 – 4	Ø 18,4 – 2,9	Ø 8,6-1,6	Ø 6,4 – 1
Working distance		110 mm	275 mm	195 mm	145 mm	50 mm	35 mm

# 7 Trouble shooting

Problem	Possible causes
You can see two images	The interpupillary distance is not set correctly
	The magnifications of the eyepieces do not match
There is dirt in the visual field	There is dirt on the object being observed
	There is dirt on the eyepiece surface
The image is unclear	There is dirt on the objective surface
The focus wheels are jammed	The torque of the focus wheels is set too high
The microscope head slips down while you are viewing the object	The torque of the focus wheels is set too low
Eyes get tired easily	The dioptre adjustment is not correct
	The brightness adjustment is not correct

### 8 Service

If, after studying the user manual, you still have questions about commissioning or using the microscope, or if unforeseen problems should arise, please get in touch with your dealer. The device may only be opened by trained service engineers who have been authorised by KERN.

### 9 Disposal

The packaging is made of environmentally-friendly materials, which you can dispose of at your local recycling centre. Disposal of the storage box and device must be carried out by the operator in accordance with all national or regional laws in force in the location of use.

#### **10 Further information**

The illustrations may differ slightly from the product.

The descriptions and illustrations in this user manual are subject to change without notice. Further developments on the device may lead to these changes.



All language versions contain a non-binding translation. The original German document is the binding version.