



Trinocular HBO fluorescence microscope, 1000x, IOS N-PLAN objectives, Blue & Green filtersets

| Observation Method - | Brightfield | Yes |
|--|---|--|
| Transmitted Light | | |
| | Phase contrast (Positive type) Darkfield | As optional As optional |
| | Simple polarized light | As optional |
| | Simple polarized light | As optional |
| Observation Method - Incident Light | Fluorescence | Yes |
| Main Body | Туре | Upright |
| main body | Construction material | Aluminum die-cast |
| | Trasportation handle | Yes |
| | | 100 |
| Head | Туре | Trinocular (Siedentopf) |
| | Split ratio | 50/50 |
| | Inclination | 30° |
| | 360° rotating | Yes |
| | Interpupillary distance (mm) | 48-75 |
| | Dioptric adjustement | On left tube |
| | Fixing screw for eyepieces | Yes |
| | Tube inner diameter (mm) | 30 |
| | | |
| Eyepieces | Field number (mm) | 20 |
| | Magnification | 10x |
| | Pointer | As optional |
| | Micrometric scale | As optional |
| | Diameter of micrometer glass (mm) | 23 |
| | High eyepoint (for glass wearers) | Yes |
| | Rubber cup | Yes |
| | | |
| Nosepiece | Positions | Quintuple |
| | Reversed | Yes |
| | Bi-directional | Yes |
| | Rotation on ball bearings | Yes |
| | Objective thread | RMS |
| | | |
| Objectives | Optical system | ∞ |
| | Anti-fungus treatment | Yes |
| | Parfocal distance (mm) | 45 |
| | Standard magnifications | 40x-1000x |
| | Туре | IOS N-PLAN |
| | | 4x/0.10, W.D. 16.8 mm |
| | | 10x/0.25, W.D. 5.8 mm |
| | | 20x/0.40, W.D. 5.1 mm |
| | | |
| | | 40x/0.65, W.D. 0.43 mm |
| | | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), |
| | | 40x/0.65, W.D. 0.43 mm |
| Stage | Tupo | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm |
| Stage | Type Dimensions (mm) | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer |
| Stage | Dimensions (mm) | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 |
| Stage | Dimensions (mm) Moving mechanism | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 Rackless |
| Stage | Dimensions (mm) Moving mechanism Moving range (mm) | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 Rackless 78x54 |
| Stage | Dimensions (mm) Moving mechanism Moving range (mm) Material | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 Rackless 78x54 Anti-scratch painting |
| Stage | Dimensions (mm) Moving mechanism Moving range (mm) Material Specimen holder | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 Rackless 78x54 Anti-scratch painting Yes |
| Stage | Dimensions (mm) Moving mechanism Moving range (mm) Material Specimen holder Slide number | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 Rackless 78x54 Anti-scratch painting Yes 2 |
| Stage | Dimensions (mm) Moving mechanism Moving range (mm) Material Specimen holder | 40x/0.65, W.D. 0.43 mm 100x/1.25 (Oil/Water), W.D. 0.13 mm Double layer 233x147 Rackless 78x54 Anti-scratch painting Yes |

| Condenser - Single | Туре | Abbe |
|------------------------------|--|---|
| Position | Removable | Yes |
| | Numerical aperture (N.A.) | 1.25 |
| | Magnification scale for simplified positioning | Yes |
| | Diaphragms | Iris |
| | Centrable | Yes |
| | Focusable | By rack and pinion |
| | | |
| Focusing System | Туре | Coaxial coarse & fine |
| J J J J J J J J J J | Coarse total travel (mm) | 25 |
| | Fine total travel (per single rotation) (mm) | 0.2 |
| | Fine graduations | 100 |
| | Fine resolution (µm) | 2 |
| | Upper stop to prevent contact | Yes |
| | Adjustable tension | Yes |
| | Flat knob for ergonomy | Yes |
| | | Tes |
| Transmitted | Kohler illumination | Fixed (full as optional) |
| Illumination | | X-LED |
| indimitation | | |
| | X-LED type | X-LED3 |
| | Light source power (W) | 3.6 Manual |
| | Brightness control | Manual |
| | Lifetime (hours) | > 65,000 |
| | Temperature (K) | 6,300 |
| | Max. required power (W) | 6 |
| | _ | - |
| Power Supply for | Туре | External |
| Transmitted | Microscope connector | Jack, 2.1 mm |
| Illumination | Power plug type | Multi-plug (EU, UK, US) |
| | Input voltage | 100/240 Vac, 50/60 Hz |
| | Output voltage | 6 Vdc 2.5 A |
| | | |
| Accessories Included | Dust cover | Yes |
| | Immersion oil (10ml) | Yes |
| | Tension adjustment tool | Yes |
| | Allen wrench | Yes |
| | User Manual | Digital version (downloadable) |
| | | |
| Additional Information | | Heating stage (as optional). External rechargeable battery pack (as optional). |
| | | External rechargeable battery pack (as optional). |
| Product Dimensions | Height (mm) | 485 |
| | Width (mm) | 235 |
| | Depth (mm) | 530 |
| | | 000 |
| Product Weight | <i>a</i> > | 12.4 |
| | (kg) | (including FLUO power supply) |
| | | |
| Fluorescence | Number of positions | 3 |
| Attachment | Blue filter set (included) | Excitation: 460 - 490 nm; Dichroic: 505 nm; |
| | | Emission: 515LP nm |
| | Green filter set (included) | Excitation: 510 - 550 nm; Dichroic: 570 nm; |
| | | Emission: 590LP nm Excitation: 25 mm diam.; |
| | Filter dimensions | Dichroic: $36 \text{ mm x} 25 \text{ mm};$ |
| | | Emission: 25 mm diam. |
| | Filter set selection | Manual |
| | Shutter | Yes |
| | | |
| Fluorescence Light | Light source | НВО |
| Source | Light source power (W) | 100 |
| | Lifetime (hours) | 300 |
| | | |
| Fluorescence Power Supply | Туре | External |
| | Power plug type | Schuko |
| | Input voltage | 100/240 Vac, 50/60 Hz |
| | Max. power required (W) | 130 |
| | LED indicators | Yes |
| | | |



Wolf Laboratories Limited

www.wolflabs.co.uk

Tel: 01759 301142

Fax:01759 301143

sales@wolflabs.co.uk



Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.





