Rev. 00



Model: **B-151R** 

Typology:

## **EDUCATIONAL MICROSCOPE**

## **Description:**

Educational microscope for routine applications, with rechargeable battery.

Dye-cast frame, with high stability and ergonomy, for transmitted light observation.

Models	<b>B-151</b> : monocular with fixed table; objectives 4x, 10x, 40x
Illumination	Light source type: white LED; light intensity control using a knob on left side of the frame.  LED power: 1W.  Color temperature: 6300K  LED average lifetime approx.: 50.000h  External power supply 100/240Vac, 50/60Hz, output: 6Vdc 2,5A.  Internal NiMH rechargeable battery (up to 15 hours of use from a single charge, led at medium intensity).
Observation Modes	Brightfield
Focusing	Coaxial coarse and fine focusing mechanism (graduated, 0.002mm) with upper stop, to prevent the contact between objective and specimen.  Adjustable tension of coarse focusing knob.
Stage	Plain stage 125X120mm.
Nosepiece	Quadruple revolving nosepiece, rotation on ball bearings.
Head	Monocular observation head, inclined 30° and rotatable 360°.
Eyepieces	Widefield eyepieces WF10X/18 with field number 18mm.
Objectives	Achromatic objectives with mechanical tube lenght 160 mm, made by following objectives:  -) Achromatic DIN 4X, A.N. 0.10, W.D. 18 mm -) Achromatic DIN 10X, A.N. 0.25, W.D. 7 mm -) Achromatic DIN 40X, A.N. 0.65, W.D. 0,53 mm All objectives are treated with an anti-fungus treatment.
Condenser	Abbe condenser N.A. 0.65 precentered.
Dimensions	HEIGHT: 350 mm  WIDTH: 170 mm  DEPTH: 210 mm  WEIGHT: 4 kg
Accessories	Instruction manual and dust cover included.



## **Wolf Laboratories Limited**

www.wolflabs.co.uk

Tel: 01759 301142

Fax:01759 301143

sales@wolflabs.co.uk







Use the above details to contact us if this literature doesn't answer all your questions.

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.





