

UVP Translinker

The UVP Translinker combines the shortwave UV CL-1000 Crosslinker and the M-20 Transilluminator. The transilluminator is built onto the crosslinker cabinet.

- On/Off controls for the transilluminator are located on the crosslinker panel
- The transilluminator uses midrange 302 nm UV and a 20 x 20 cm filter



| Order Number | | Description |
|--------------|--------------|---|
| 115 V | 230 V | UVP Crosslinkers, UVP Translinkers and Replacement Tubes |
| 95-0174-01 | 95-0174-02 | CL-1000 Crosslinker |
| 95-0228-01 | 95-0228-02 | CL-1000L Crosslinker |
| 95-0230-01 | 95-0230-02 | CL-1000M Crosslinker |
| 95-0339-01 | 95-0339-02 | CX-2000 Crosslinker |
| 95-0300-01 | 95-0300-02 | TL-2000 Translinker |
| | 34-0007-01 | Replacement Tube (CL-1000, CX-2000, TL-2000), 8-watt, 254 nm UV |
| | 34-0006-01 | Replacement Tube (CL-1000L), 8-watt, 365 nm UV |
| | 34-0042-01 | Replacement Tube (CL-1000M, TL-2000), 8-watt, 302 nm UV |

UVP Incubator

This UV sterilizing incubator provides precise temperature control and uniformity throughout the chamber for incubation of biological assays, fungal, bacterial cultures, eggs, and other samples.



- Microprocessor-controlled incubation temperature
- Heating element and fan generate uniform temperature throughout the chamber
- Decontaminate the chamber between growth experiments with the overhead shortwave 254 nm UV sterilization
- Two ventilated shelves included; remove the shelves to stack up to 150 (10 cm) plates
- Temperature Ambient to 68°C, Accuracy: +0.5°C
Uniformity: +0.1°C at 37°C, Humidity: ~80%

| Order Number | | Description |
|--------------|--------------|---|
| 115 V | 230 V | UVP Incubator and Replacement Tube |
| 95-0358-01 | 95-0358-02 | UV Benchtop Incubator (SI-950) |
| | 34-0007-01 | Replacement Tube, 8-watt, 254 nm UV |



WolfLabs

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

www.wolflabs.co.uk

Tel : 01759 301142

Fax : 01759 301143

sales@wolflabs.co.uk

Please contact us if this literature doesn't answer all your questions.