

Model CB 170 | CO₂ incubators with hot air sterilization and heat sterilizable CO₂ sensor

The BINDER CB series CO₂ incubator is the premium class among the CO₂ incubators. It is suitable for all sensitive incubation applications and ensures optimal cell growth. A BINDER CO₂ incubator of the CB series is up to even complex cultivation experiments or individual environments.

BENEFITS

- Maximum usable space with minimal footprint and excellent ease of use
- Minimum risk of contamination thanks to uncompromising hot-air sterilization
- Fanless, easy-to-clean hygienic interior
- Stable pH values thanks to drift-free CO₂ IR sensor technology
- Low evaporation of the medium due to high levels of humidity and quick recovery time



Model CB 170



Model CB 170

MAIN FEATURES

- Temperature range: ambient temperature plus 4 °C to 60 °C
- Humidity range: 90 to 95% RH
- CO₂ control range: 0 to 20 vol% CO₂
- O₂ control range: 0.2 to 20 vol% O₂ for versions with O₂ control
- Alternative O₂ control range: 10 to 95 vol% O₂, option for versions with O₂ control
- Tried-and-tested anti-contamination concept with 180 °C hot air sterilization
- Hot-air sterilizable CO₂ sensor with single-beam infrared technology
- CO₂ gas-mixing jet with Venturi effect
- Intuitive touchscreen controller
- Internal data recording, can be read out in open format via USB interface
- Optimized double-pan humidification system with condensation protection
- 170 liters of extra space in the interior while maintaining the same footprint
- Stacking height of just 1.80 m for easy use of the top unit
- Fail-safe CO₂ automatic diagnosis system for protecting the pH of cell cultures
- DuoDoor™ system: Unit door and glass door can be opened with one handle
- Lockable door handle
- Computer interface: Ethernet
- Zero-voltage alarm contact
- Tightly sealed inner door made of tempered safety glass
- 3 perforated shelves, stainless steel
- Units are stackable with stacking adapter

ORDERING INFORMATION

Interior volume [L]	Voltage	Option model	Version	Art.-No.
Model CB 170				
170	200...240 V 1~ ph 50/60 Hz	Standard	CB170-230V	9040-0131
		with O ₂ control	CB170-230V-O	9040-0132
		With active humidification	CB170-230V-F	9040-0133
		With O ₂ control 0.2–20% and active humidification	CB170-230V-OF	9040-0134
	100...120 V 1~ ph 50/60 Hz	Standard	CB170UL-120V	9040-0139
		with O ₂ control	CB170UL-120V-O	9040-0140
		With active humidification	CB170UL-120V-F	9040-0141
		With O ₂ control 0.2–20% and active humidification	CB170UL-120V-OF	9040-0142

TECHNICAL DATA (Version 1-4 from 8)

Description	CB170-230V ¹	CB170-230V-O ¹	CB170-230V-F ¹	CB170-230V-OF ¹
Article Number	9040-0131	9040-0132	9040-0133	9040-0134
Performance Data Temperature				
Temperature range 4 °C above ambient temperature to [°C]	60			
Temperature range 6 °C above ambient temperature to [°C]		60	60	60
Temperature variation at 37 °C [± K]	0.3	0.4	0.4	0.4
Temperature fluctuation at 37 °C [± K]	0.1	0.1	0.1	0.1
Recovery time after 30 seconds door open at 37 °C [min]	6	6	6	6
Performance Data Climate				
Humidity range [% RH]	90...95	90...95	90...95	50...95
CO ₂ -Data				
CO ₂ range [Vol.-% CO ₂]	0...20	0...20	0...20	0...20
CO ₂ measurement	IR	IR	IR	IR
CO ₂ Recovery time after 30 seconds door open at 5 Vol.-% CO ₂ [min]	5	5	5	5
O ₂ -Data				
Standard O ₂ control ranges: O ₂ range [Vol.-% O ₂]		0.2...20		0.2...20
O ₂ control ranges with option: O ₂ range [Vol.-% O ₂]		10...95		10...95
O ₂ Recovery time after 30 seconds door open at 5 Vol.-% O ₂ [min]		10		12
Electrical data				
Rated Voltage [V]	200...240	200...240	200...240	200...240
Power frequency [Hz]	50/60	50/60	50/60	50/60
Nominal power [kW]	1.4	1.4	1.4	1.4
Unit fuse [A]	10	10	10	10
Phase (Nominal voltage)	1~	1~	1~	1~
Internal Dimensions				
Depth [mm]	505	505	505	505
Height [mm]	600	600	600	600
Width [mm]	560	560	560	560

1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

Description	CB170-230V ¹	CB170-230V-O ¹	CB170-230V-F ¹	CB170-230V-OF ¹
Article Number	9040-0131	9040-0132	9040-0133	9040-0134
Housing dimensions not incl. fittings and connections				
Width net [mm]	680	680	680	680
Height net [mm]	870	870	870	870
Depth net [mm]	715	715	715	715
Measures				
Wall clearance back [mm]	100	100	100	100
Wall clearance sidewise [mm]	50	50	50	50
Interior volume [L]	170	170	170	170
Net weight of the unit (empty) [kg]	95	95	95	95
permitted load [kg]	30	30	30	30
Load per rack [kg]	10	10	10	10
Doors				
Inner doors	1	1	1	1
Unit doors	1	1	1	1
Environment-specific data				
Energy consumption at 37 °C [Wh/h]	55	55	55	55
Sound-pressure level [dB(A)]	41	41	41	41
Fixtures				
Number of shelves (std./max.)	3/6	3/6	3/6	3/6

1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

TECHNICAL DATA (Version 5-8 from 8)


Description	CB170UL-120V ¹	CB170UL-120V-O ¹	CB170UL-120V-F ¹	CB170UL-120V-OF ¹
Article Number	9040-0139	9040-0140	9040-0141	9040-0142
Performance Data Temperature				
Temperature range 4 °C above ambient temperature to [°C]	60			
Temperature range 6 °C above ambient temperature to [°C]		60	60	60
Temperature variation at 37 °C [± K]	0.3	0.4	0.4	0.4
Temperature fluctuation at 37 °C [± K]	0.1	0.1	0.1	0.1
Recovery time after 30 seconds door open at 37 °C [min]	6	6	6	6
Performance Data Climate				
Humidity range [% RH]	90...95	90...95	50...95	50...95
CO ₂ -Data				
CO ₂ range [Vol.-% CO ₂]	0...20	0...20	0...20	0...20
CO ₂ measurement	IR	IR	IR	IR
CO ₂ Recovery time after 30 seconds door open at 5 Vol.-% CO ₂ [min]	5	5	5	5

1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

Description	CB170UL-120V ¹	CB170UL-120V-O ¹	CB170UL-120V-F ¹	CB170UL-120V-OF ¹
Article Number	9040-0139	9040-0140	9040-0141	9040-0142
O₂-Data				
Standard O ₂ control ranges: O ₂ range [Vol.-% O ₂]		0.2...20		0.2...20
O ₂ control ranges with option: O ₂ range [Vol.-% O ₂]		10...95		10...95
O ₂ Recovery time after 30 seconds door open at 5 Vol.-% O ₂ [min]		10		12
Electrical data				
Rated Voltage [V]	100...120	100...120	100...120	100...120
Power frequency [Hz]	50/60	50/60	50/60	50/60
Nominal power [kW]	1.4	1.4	1.4	1.4
Unit fuse [A]	16	16	16	16
Phase (Nominal voltage)	1~	1~	1~	1~
Internal Dimensions				
Depth [mm]	505	505	505	505
Height [mm]	600	600	600	600
Width [mm]	560	560	560	560
Housing dimensions not incl. fittings and connections				
Width net [mm]	680	680	680	680
Height net [mm]	870	870	870	870
Depth net [mm]	715	715	715	715
Measures				
Wall clearance back [mm]	100	100	100	100
Wall clearance sidewise [mm]	50	50	50	50
Interior volume [L]	170	170	170	170
Net weight of the unit (empty) [kg]	95	95	95	95
permitted load [kg]	30	30	30	30
Load per rack [kg]	10	10	10	10
Doors				
Inner doors	1	1	1	1
Unit doors	1	1	1	1
Environment-specific data				
Energy consumption at 37 °C [Wh/h]	55	55	55	55
Sound-pressure level [dB(A)]	41	41	41	41
Fixtures				
Number of shelves (std./max.)	3/6	3/6	3/6	3/6

1 All technical data is specified for unloaded units with standard equipment at an ambient temperature of +22 °C ±3 °C and a power supply voltage fluctuation of ±10%. The temperature data is determined to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

DIMENSIONS incl. fittings and connections [mm]



OPTIONS AND ACCESSORIES

Designation	Description	*	Art.-No.
Access port with silicone plug	back		
	30 mm	01	8012-1390
	left		
	30 mm	01	8012-1391
	right		
	30 mm	01	8012-1392
Analog output 4-20 mA	for temperature and CO ₂ values (outputs not adjustable)	02	8012-1839
	for temperature and CO ₂ values (outputs not adjustable), for versions with O ₂ control	02	8012-1840
APT-COM™ 4 BASIC-Edition	for simple logging and documentation requirements with up to 5 networked units. version 4, BASIC edition	19	9053-0039
APT-COM™ 4 GLP-Edition	for working under GLP-compliant conditions. Measured values are documented in a tamper-proof way in line with the requirements of FDA Regulation 21 CFR 11. version 4, GLP edition	19	9053-0042
APT-COM™ 4 PROFESSIONAL-Edition	convenient unit and user management built on the BASIC edition. Suitable for networking up to 100 units. version 4, PROFESSIONAL edition	19	9053-0040
Base	with casters, height-adjustable, with unit feet retainer	–	9051-0028
Calibration certificate O ₂	for versions with O ₂ control; a measuring point in center of chamber at a specified O ₂ percentage value	–	8012-0229
Calibration certificate, expanded	for temperature; for extending the measurement in center of chamber to include another test temperature	–	8012-1113
Calibration certificate, temperature	for temperature, measurement in center of chamber at specified temperature	–	8012-1132
	temperature measurement incl. certificate and 27 measuring points at specified temperature	–	8012-1592
	temperature measurement incl. certificate, 15- 18 measuring points at specified temperature	–	8012-1571
	temperature measurement incl. certificate, 9 measuring points at specified temperature	–	8012-1550
Calibration certificate, temperature and CO ₂	for temperature and CO ₂ , temperature measurement in center of chamber, CO ₂ measurement performed using test gas at 37 °C and 5 % CO ₂	–	8012-1235
CELLROLL set	modular, expandable roller bottle system for cell cultivation, 4 roller bottles	–	8012-1885
Copper interior fittings	Set consisting of 3 shelves (all-copper) and a water pan (copper-plated), more information.	–	8012-1918
Door hinged on the left	Unit door and standard glass inner door with door hinged on the left	–	8012-1881
Electric access port	8-pin, for low voltage with LEMO socket (coverable) and LEMO plug (max. 24 V – 2 A)	–	8012-1848
Gas cylinder connection set	for CO ₂ , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	–	8012-0014
	for N ₂ , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	–	8012-0016
	for O ₂ , consisting of a gas tank pressure regulator with connection parts and 5-meter hose	–	8012-0015
Gas tank changer	external, BINDER Gas Supply Service, for connecting 2 gas tanks (CO ₂ , N ₂ or O ₂), with audible and visual alarms, as well as zero-voltage alarm output	–	8012-0408
Inner door, divided	Internal door divided into 6 parts		
	door hinged on the left (incl. option door hinged on the left), 120 V option model	–	8012-1889
	door hinged on the left (incl. option door hinged on the left), 230 V option model	–	8012-1880
	door hinged on the right	–	8012-1844
O ₂ control range	alternative to the standard range, for controlling O ₂ concentrations between 10 and 95 vol. %, only in conjunction with option models with O ₂ control	–	9418-0207
pH-neutral detergent	concentrated, for gentle remove of residual contaminants; 1 kg	–	1002-0016
Qualification folder IQ/OQ	supporting documents for validation performed by customer, consisting of: IQ/OQ checklists, unit schematics, QM certificate in accordance with ISO 9001	15, 18, 20	8012-0876
Qualification folder IQ/OQ/PQ	supporting documents for validation performed by customer, in accordance with customer's requirement, extension of Qualification folder IQ/OQ by chapter PQ	15, 18	8012-0963
Quick sample access	viewing window (13 x 25 cm) in the inner door for quick access		
	door hinged on the left (incl. option door hinged on the left), 120 V option model	–	8012-1890
	door hinged on the left (incl. option door hinged on the left), 230 V option model	–	8012-1884
	door hinged on the right	–	8012-1883
Rubber pads	set anti-slip feet	–	8012-0702

* Notes › See last page

Designation	Description	*	Art.-No.
Shaker „Professional“	Shaker and controller equipped with LEMO connectors, exclusively for CB with electric access port, 100 – 240 V, 50/60 Hz	–	8012-1772
Shaker „Universal“	Shaker and controller connected via ribbon cable, universally applicable, 100 – 240 V, 50/60 Hz	–	8012-1771
	Stainless steel	–	6004-0137
Shelf, perforated	for divided inner door		
	Stainless steel, divided	–	8012-1888
Stacking adapter, flat	for thermally decoupled stacking of two identical units	–	9051-0035
Stacking frame	stable, vibration-free stacking frame on casters with stop brake for safe stacking of two CB series units	–	9051-0020

* Notes › See last page

SERVICES

Designation	Description	*	Art.-No.
Installation services			
Installation	and set up of unit at operating location, connect to existing connections	13, 18	DL10-0100
Instruction	unit function instructions for operation and programming of the controller	18	DL10-0500
Preventive maintenance			
Preventive maintenance	Executive of equipment inspection according to maintenance plan	14, 18	DL20-0400
Calibration services			
Calibration CO ₂	including certificate, CO ₂ measurement with analyzed test gas at 5 % or at specified value	14, 16, 17, 18	DL30-0401
Calibration O ₂	calibration of O ₂ , including certificate (one measuring point in center of chamber at specified O ₂ percentage)	14, 16, 17, 18	DL30-0402
Calibration temperature incl. certificate	including certificate, one measuring point in center of chamber at specified temperature	14, 16, 17, 18	DL30-0101
Temperature measurement 18 measuring points	including certificate, 18 measuring points at specified temperature	14, 16, 17, 18	DL30-0118
Temperature measurement 27 measuring points	including certificate, 27 measuring points at specified temperature	14, 16, 17, 18	DL30-0127
Temperature measurement 9 measuring points	including certificate, 9 measuring points at specified temperature	14, 16, 17, 18	DL30-0109
Validation services			
Execution of IQ/OQ	in accordance with qualification folder	15, 18, 20	DL41-0200
Execution of IQ/OQ/PQ	in accordance with customer's requirement, price: on request	15, 18	DL44-0500
Qualification folder IQ/OQ	supporting documents for validation performed by customer, consisting of: IQ/OQ checklists, unit schematics, QM certificate in accordance with ISO 9001	15, 18, 20	8012-0876
Qualification folder IQ/OQ/PQ	supporting documents for validation performed by customer, in accordance with customer's requirement, extension of Qualification folder IQ/OQ by chapter PQ	15, 18	8012-0963
Warranty service			
Extension of the warranty from 2 to 3 years	beginning with the date of delivery, wearing parts are not included	–	DL02-1041
Extension of the warranty from 2 to 5 years	beginning with the date of delivery, wearing parts are not included	–	DL02-1042

* Notes › See last page



WolfLabs

Pricing on any accessories shown can be found by keying the part number into the search box on our website.

The specifications listed in this brochure are subject to change by the manufacturer and therefore cannot be guaranteed to be correct. If there are aspects of the specification that must be guaranteed, please provide these to our sales team so that details can be confirmed.

www.wolflabs.co.uk

Tel : 01759 301142

Fax : 01759 301143

sales@wolflabs.co.uk

Please contact us if this literature doesn't answer all your questions.